

I. OCT in glaucoma: interpretation, progression, and management

- A. Goals
 - 1. Discuss glaucoma diagnosis and progression in terms of OCT
 - 2. Review strengths and weaknesses of OCT for glaucoma
 - 3. Summarize research about OCT related to diagnosis and progression
 - 4. Provide tips for better OCT analysis in clinical practice

II. OCT Overview

- A. Available OCTs
 - 1. Zeiss Cirrus, Heidelberg Spectralis, Topcon, RTVue, Zeiss Stratus, etc
- B. Spectral domain (SD) vs. time domain (TD)
 - 1. Spectral also known as Fourier Domain
 - 2. Similarities/differences in databases
 - a) **Can't directly compare progression between instruments**
 - 3. SD-OCT advantages:
 - a) Higher resolution, decreased scanning time
 - b) Better repeatability of RNFL measurements
 - c) More data – peripapillary scans with RNFL thickness maps, macular and ganglion cell (GC) analysis, etc
 - d) Better for diagnosis and progression vs TD, HRT, GDx
- C. Measurement Boundaries
 - 1. Retina
 - a) Inner boundary – ILM for all
 - b) Outer boundary varies: Stratus – photoreceptor inner/outer segment interface, Topcon/Copernicus – inner RPE, Cirrus – outer RPE, RTVue – external RPE, Spectralis – **Bruch's (Giani, Grover 2010)**
 - 2. Disc margin
 - a) Stratus reference – RPE/choriocapillaris plus 150 um above RPE
 - b) SD-OCT – **Bruch's membrane is reference**
- D. Properties of selected OCTs
 - 1. Zeiss Stratus (Time Domain)
 - a) 3.46 mm scan around ONH
 - b) 400 axial scans/second
 - c) 7-8um axial resolution
 - d) Motion artifact/centration are difficult
 - e) Widely used worldwide
 - f) Many clinical trials
 - 2. Zeiss Cirrus (Spectral Domain or SD)
 - a) 6x6mm² cube over ONH
 - b) 200 B-Scans with 200 A-Scans each
 - c) 27,000 axial scans/second
 - d) 3.46mm circle over ONH for clock hour positions
 - e) 5 micron axial resolution
 - f) Cup/disc dimension measurements

- g) Vert and horiz c/d, disc area, cup volume, etc
- h) Thickness deviation maps: 50x50 superpixels
- i) 92.1-98.3% sensitive for glaucoma detection (Leung 2010)
- 3. Heidelberg Spectralis (SD)
 - a) 6 consecutive circular B-Scans
 - b) 40,000 axial scans/second
 - c) 12mm scan diameter
 - d) 3.45-3.6mm scan around ONH
 - e) Scan circle varies with axial length
 - f) 3.9 micron axial resolution
- 4. Optovue RTVue-100 (SD)
 - a) 5 micron axial resolution
 - b) 26,000 axial scans/second
 - c) 13 circular scans of 1.3-4.9mm diameter
 - d) Eye tracking
- E. Additional features of some SD-OCT
 - 1. FastTrack (Cirrus), FoDi (Spectralis), GPA – help follow over time
 - 2. Ability to manually adjust scans

III. Interpretation and Influential Factors

- A. Guide to Interpreting Spectral Domain Optical Coherence Tomography by Bruno Lumbroso, MD and Marco Rispoli, MD
- B. Normative databases
 - 1. Stratus
 - a) 328 subjects
 - b) 48% male, 52% female
 - c) Mean age 47.4 +/- 15.8 yrs, range 18-85
 - d) Rx: -11.75 to +6.75, mean -0.54
 - e) 63% Caucasian, 24% Hispanic, 8% African American, 11% Asian
 - f) No eye surgery except cataract (9 pts), no ocular disease, IOP <22, normal and reliable VF, normal ONH, BCVA >20/32
 - 2. Cirrus
 - a) 284 subjects
 - b) 47% male, 53% female
 - c) Age range 19-84
 - d) Rx: -12 to +8
 - e) 43% Caucasian, 18% African American, 12% Hispanic, 1% Indian, 6% mixed
 - f) All normal subjects
 - 3. Spectralis
 - a) 201 subjects, all Caucasian
 - b) 55% male, 45% female
 - c) Mean age 48.2 +/- 14.5 yrs, Range 18-78
 - d) Only 1 pt <20 and only 13 pts >70
 - e) Rx: -7 to +5
 - f) No glaucoma, normal IOP, normal VF, normal optic nerve, etc
 - 4. RTVue

- a) 861 subjects (largest of SD-OCT), various ethnicities
 - b) Mean age 50 +/- 15.5 yrs, Range 19-82
 - c) Rx: -8 to +8 sphere, -2 to +2 cylinder
 - d) No glaucoma, normal IOP, normal VF, normal optic nerve, etc
- C. Red-Green disease
- 1. Color code determined by the database of the instrument
 - a) Based on probability of that population only
 - 2. 15-36% of OCTs for glaucoma may contain artifacts that influence red-green analysis
- D. Average RNFL Thickness measurements
- 1. Stratus
 - a) Average 94-100um in Caucasian/Japanese
 - b) Up to 132.7um in Hispanics
 - c) OCT1 and OCT2: 86-153um
 - 2. Spectralis
 - a) Thicker ppRNFL vs. Cirrus in same eyes
 - b) Average 89-97.3um +/- 9.6-15.87um
 - c) African American 99.2um, Caucasian 96um
 - 3. Cirrus
 - a) Average 84-94um +/- 13.68
 - b) Superior quad up to 122um, inferior quad up to 127um
 - 4. RTVue
 - a) Average 107.9 +/- 10um
 - b) African American 107um, Caucasian 102um
 - 5. Topcon
 - a) Average 102um
- E. ISNT Rule
- 1. RNFL thickness on OCT usually matches neuroretinal rim appearance, but:
 - a) Only 42% of normals follow RNFL ISNT with Spectralis
 - b) Only 79% of normals with HRT
 - c) Only 28% of glaucoma pts with HRT
 - 2. IST rule
 - a) 47.1% of normals follow ISNT, but 58.7% follow IST with Stratus
 - b) 25.9% of normals follow ISNT, but 70.4% follow IST with HRT
 - (1) Neither is very useful clinically
- F. Media/PVD effect scan quality
- 1. Can significantly reduce quality of scans
 - 2. PCIOLs do not seem to have significant effect (Kim 2013)
- G. ONH size/Disc area
- 1. Larger ONH means OCT scan is closer to ONH
 - a) RNFL thickness decreases as measurement diameter increases
 - b) Overestimates RNFL in some studies but not others
 - 2. Thicker RNFL measurements in larger ONH
 - a) 3.3um per 1mm² (Budenz 2007)
 - 3. RNFL Thickness correlates with disc area (Hirasawa 2010, Japanese)
 - 4. No association between ONH size and RNFL thickness (AIGS 2012)
 - 5. Disc area measurements

- a) Cirrus
 - (1) Small: $<1.66\text{mm}^2$
 - (2) Medium: $1.63-1.97\text{mm}^2$
 - (3) Large: $>1.97\text{mm}^2$
 - (4) Only 5% of eyes in normal database were $<1.33\text{mm}^2$ or $>2.5\text{mm}^2$ with Cirrus
- b) Stratus
 - (1) 2.26mm^2 mean disc area
- c) RTVue
 - (1) Range $1.86-2.1\text{mm}^2$

H. PPA

1. Present in 15% of normals but 62-84% of glaucoma patients
2. Disc size variations between instruments
 - a) Stratus overestimates disc size in glaucoma patients and controls
 - b) Cirrus performs well compared to clinical disc evaluation

I. ERM

1. Most common cause of artifact in RNFL determination

J. Peripapillary retinoschisis

1. Can cause false impression of thick RNFL
2. After resolution, can simulate glaucomatous thinning

K. Axial length

1. Some studies found no correlation with axial length and RNFL thickness (Hirasawa 2010)
2. Others show total RNFL thickness decreases with increased axial length ($2.2\text{um}/1\text{mm}$ in Stratus)
3. If temporal quadrant is thick, superior and inferior thinning could be due to refractive error (Alasil 2012)
 - a) 60.3% supernormal sectors in Japanese myopes, mostly temporal, indicating false positive (Yamashita 2014)
4. Be cautious of thinning in myopic Caucasians
5. Stratus database may be inaccurate (Vernon 2008)

L. ONH distance to foveola

1. High myopia: RNFL bundles converge causing abnormalities (Leung 2012)
2. Temporal or nasal deviated RNFL plot can over diagnose glaucoma
3. Spectralis FoDi example

M. Other Factors

1. Rx: RNFL thinner by $1.2\text{um}/\text{diopter}$ of myopia
2. Race: RNFL decreases from Hispanics>Asians> African Americans>Caucasians
3. Patients with FOHx of glaucoma have thinner RNFL and GCC than normals (Rolle 2014)

N. Interocular symmetry

1. Increasing age is not associated with increased RNFL asymmetry
2. Cirrus: $>9\text{um}$ difference may be indicative of early glaucoma
3. Spectralis: 6.6x greater asymmetry in glaucoma vs. normal
 - a) Difference of 6um for RNFL global average had high sensitivity and specificity to detect POAG
 - b) Use absolute RNFL thickness and RNFL asymmetry analyses

(1) **Asymmetry differences aren't color coded** (yet)

4. Macular asymmetry has also proven sensitive and specific (Sullivan-Mee)

O. Case examples to illustrate points throughout the presentation

IV. Glaucoma Diagnosis and Progression

A. Utility

1. RNFL loss precedes VF loss by 6 years in 60% of eyes (Sommer 1991)
2. In OHTS, HRT showed glaucomatous change 8 years before VF defects
3. 17% RNFL loss before VF detection (Wollstein 2012)
4. Progressive optic disc changes may not correspond to RNFL thinning in the same eyes with glaucoma progression

B. Glaucoma Detection

1. Both TD and SD have high sensitivity and specificity for glaucoma when >1 clock hour is <5% level (red)
2. Both TD and SD may be inadequate in detecting preperimetric RNFL defects
 - a) Stratus has difficulty determining severity of glaucoma (Smith 2014)
3. Cirrus can discriminate mild glaucoma from normal based on ONH parameters
4. RNFL parameters:
 - a) Average RNFL thickness*
 - b) RNFL thickness at 7 o'clock (inf-temp, OD reference)
 - (1) 3,4,9 are most variable
 - c) RNFL thickness inferior quadrant
 - d) Global, sup-temporal and inf-temporal (Spectralis)
5. ONH parameters:
 - a) Vertical rim thickness (VRT)
 - b) Rim area
 - c) Vertical C/D (VCDR)
6. Additional helpful information:
 - a) Cirrus – RNFL thickness map and deviation-from-normal map
 - (1) Yellow if exceeds test-retest variability once
 - (2) Red if exceeds on consecutive visits
 - b) Stratus – TSNIT
7. Case Examples

C. Progression considerations

1. Variable nature of glaucoma
2. Event-based vs. trend-based analyses
 - a) Event: difference between baseline and follow-up measurements exceeds test-retest variability limit
 - b) Trend: linear regression analysis of a parameter (i.e. average RNFL) over time showing negative slope
3. Changing technology – longitudinal follow-up difficulties
 - a) Dr. George Spaeth and others prefer disc photos as gold standard
4. Instrument variability
5. No consensus on limit of RNFL thinning that equals progression; no reference standard

a) In patients without VF loss, it is hard to determine if OCT structural changes are false positives or if they are structural change before functional change

D. Progression: Various methods

1. Average RNFL thickness may be better than sector analysis with lower inter-test variation
2. Significant negative trend in average RNFL thickness with time?
 - a) -1.52um to -5.03um/year for Cirrus
 - b) -2.22um to -7.60um/year for Stratus
3. >1 clock hr at the <5% level?
4. **1 clock hr at <5% and overall 'borderline' or 'outside normal'?**

E. Reliability and reproducibility

1. Inter-visit repeatability is good for most SD-OCT
2. Signal strength: 7 or greater desired
3. Dilation: may not effect repeatability
4. Variability vs. progression?
5. Test vs. re-test fluctuations
 - a) Stratus: ~4-10um per quadrant
 - (1) Longitudinal changes up to 11.7um occur
 - (2) Be suspicious of changes over 10um
 - b) Cirrus: >4-6um between visits is suspicious
 - (1) 2 superpixels could show progression
 - c) Spectralis: 5-14um intra- and inter-visit variation
 - (1) Clinically appears to have very low fluctuation
 - (2) -2.12um/yr in progressing pts vs. -1.18um/yr in stable pts
 - d) RTVue: 100% of RNFLT of normals remained normal over 4 years

F. Case examples to illustrate inter-test variation

G. Recommendations

1. Repeat OCTs before making treatment decisions
 - a) 41-56% of abnormal scans were not duplicated on f/u exams
2. Consider abnormal if 2 of 3 RNFL or GCIPL scans are borderline or ONL

V. Types of RNFL loss and analysis

A. Patterns of RNFL loss on OCT

1. Diagnostic criteria: More than 1 clock hr at <5% level (yellow)
 - a) 90.5-96.6% sensitive on Cirrus
 - b) 85.7-91.4% sensitive on Stratus
2. Diagnostic criteria: Average RNFL thickness at <1% (red)
 - a) 44.4-72.4% sensitive on Cirrus
 - b) 33.3-60.3% sensitive on Stratus
3. RNFL deviation map was better than peripapillary RNFL measurements

B. Types of RNFL changes (Leung 2012 – Cirrus with GPA)

1. Widening of RNFL defect (85.7%)
 - a) Angular width of defects can be a useful alternative for RNFL average thickness
2. Deepening of RNFL defect

- a) RNFL defect depth percentage index (RDPI) (Suh 2014)
 - (1) New parameter on Cirrus RNFL thickness deviation map
 - (2) Better discrimination than cpRNFL between mild and moderate RNFL defects, but not moderate and severe
 - 3. Development of new RNFL defect (17.9%)
 - 4. Inferotemporal meridian is most common in glaucoma
 - a) ONH parameters, RNFL thickness and GCIPL thickness vary according to optic disc morphology and initial area of glaucomatous damage (Shin 2014)
 - 5. Other optic neuropathies can cause RNFL thinning, but patterns are different
 - a) Compressive optic neuropathy has nasal and temporal thinning
- C. Rates of Change and Age-Related RNFL Loss
 - 1. Average rate: -0.10 to -0.52um /yr (1.5-2mm/decade)
 - 2. Influenced by baseline thickness
 - a) Greater baseline thickness = faster rate of change
 - 3. No significant change in nasal and temporal quadrants with age
 - 4. Rates between normal and glaucoma pts vary:
 - a) -0.17 to -0.86 um/yr is normal
 - b) -2.54 um/yr is significant (outside 95% confidence)
 - 5. **Highest progression rates in 6 o'clock sector of cRNFL (-2.35 um/yr) and inferior outer sector of macula (-2.879 um/yr)** (Na 2014)
 - a) Perimetric glaucoma had higher rates than pre-perimetric
 - 6. Rate of global RNFL loss was more than 2x as fast in those who developed VF defects (Miki 2014) (-2.02 um/yr vs. -0.82 um/yr)
 - a) 1 um/yr faster RNFL loss = 2.05x risk of developing VF defect
- D. Correlation between RNFL changes and VF defects
 - 1. Low agreement for progression on both VF and OCT; 0.9 to 46.4% (Leung 2011, 2012)
 - 2. OCT accuracy is effected by severity
 - a) Better in more severe glaucoma (Leite 2010)
 - 3. Faster rate of RNFL thinning by OCT than VF loss (Wollstein 2005)
 - 4. Eyes that progress on VF have faster rate of RNFL loss on OCT (Grewal 2012)
- E. Progression analysis software for OCT (GPA)
 - 1. Cirrus: GPA available for OCT or HVF or combined analysis for both
 - 2. Pros: OCT GPA on Cirrus is useful to judge progression when VF defect is mild
 - 3. Cons: Agreement between OCT GPA and disc photos or VF analysis can be poor
 - 4. Example cases

VI. Macular OCT analyses for glaucoma

- A. Macular OCT
 - 1. Utility in advanced glaucoma due to papillomacular bundle preservation
 - 2. Macular OCT may be better for progression in moderate and severe glaucoma
 - a) Other studies show RNFL average thickness is still better (Grewal 2013)

b) Macular thickness rate of change was higher than RNFL thickness change

(1) -2.43um/yr to -0.98um/yr, foveal rate highest

B. Ganglion cell analysis (GCA)

1. Macular RGC complex is 1-7 cells thick: RNFL, GCL and IPL
 - a) GCIPL is less variable than RNFL and ONH
 - b) Contains 50% of retinal RGCs
 - c) Average RGC count is lower in eyes with early VF defects: 652K vs 911K (Medeiros 2013)
 - d) RGC loss of 7877 per year (Medeiros 2012)
2. RGC counts performed better than average RNFL thickness for separating glaucomatous eyes with early/minimal VF loss from healthy eyes
 - a) GCIP and GCC measurements are able to diagnose pre-perimetric glaucoma (Topcon)
 - b) Pattern deviation on SAP may underdiagnose glaucoma cases that have diffuse loss of sensitivity
 - c) However, macular RGC counts can be affected by drusen and AMD
3. GCIPL and total macular thickness (TMT) have similar sensitivity in detecting glaucoma progression, but average RNFL was better in diagnosis (Na 2012)
4. GCIPL thinning with thinner RNFL, older age, longer axial length, and males (Mwanza 2011)
5. Minimum GCIPL is best parameter for early perimetric glaucoma detection and is similar to best RNFL or ONH parameters (Mwanza 2014)
6. Macular GCA maps were useful in early glaucoma detection, but missed abnormal findings when angular distance from fovea to RNFL defect was large (Hwang 2014)
7. Macular mean sensitivity on HVF and GCIPLT showed stronger correlation with worsening of glaucoma compared to pRNFLT (Kim KE 2014)
 - a) Relationship was not established in preperimetric cases

VII. Considerations

A. Problems with progression

1. **VF testing isn't as good in early stages**
2. **OCT isn't as good in late stages**
3. Clinical trials show structural or functional changes can occur first
4. Discrepancies in literature
 - a) Only a moderate association between VF regions and RNFL thickness in glaucoma patients (Ferrerias 2008)
 - b) Agreement in progression detection between OCT of RNFL and ONH rim with VF is poor and rates vary considerably (Leung 2011)
 - c) Linear relationship exists between VF and RNFL loss (Grewal 2009)
5. Combining structural and functional tests may be the best technique (Medeiros 2012)

B. Adjunct and future technology

1. Enhanced depth imaging (EDI) – inverted image by moving SD-OCT closer to eye, can be used for choroid or lamina
2. Swept source OCT (SS-OCT) – longer wavelength (150nm) and tunable laser can image retina and choroid together
3. SD-OCT integrated with adaptive optics (AO-OCT) – corrects optical aberrations for better resolution

4. Polarization-sensitive SD-OCT – intrinsic tissue properties measured along with structure
5. Micro-OCT and ultra-high resolution OCT are coming with 1um resolution
6. These techniques are now allowing for excellent visualization of lamina cribrosa to see pores, displacement, etc.

VIII. Summary points

- A. OCT is great technology but it isn't perfect
 1. OCT is validated for glaucoma diagnosis/screening and has been shown to be highly repeatable
- B. Need to evaluate scan data and not just color codes
 1. Confounding factors, artifacts, instrument capabilities, etc
 2. Repeat OCTs before making treatment decisions
- C. Progression can be judged many ways and they do not always agree
 1. Limitations due to slow, variable nature of glaucoma
 2. Currently there is no set standard for glaucoma progression on OCT
 - a) Limited long-term follow-up data
 3. **OCT can't detect disc hemes, pallor**, etc so need to combine that information with clinical exam, disc evaluation and HVF
 4. Keep in mind limitations of normal database vs. monitoring changes in individual patients

IX. References

1. Nduaguba C, Lee RK. Glaucoma screening: Current trends, economic issues, technology, and challenges. *Curr Opinion in Ophthalmol* 2006; 17: 142-152.
2. Susanna Jr, R. Unpredictability of glaucoma progression. *Current Med Research and Opinions*, 2009; 125(9): 2167-2177.
3. Drance SD, Anderson DR, Schulzer M (For the collaborative normal-tension glaucoma study group). Risk factors for progression of visual field abnormalities in normal-tension glaucoma. *Am J Ophthalmol*, 2001; 131: 699-708.
4. Giangiacomo A, Garway-Heath D, Caprioli J. Diagnosing glaucoma progression: Current practice and promising technologies. *Curr Opinion in Ophthalmol*, 2006; 17: 153-162.
5. Wollstein et al. Optical coherence tomography longitudinal evaluation of retinal nerve fiber layer thickness in glaucoma. *Arch Ophthalmol*, 2005; 123(9): 464-470.
6. Savini G, Carbonelli M, Barboni P. Spectral-domain optical coherence tomography for the diagnosis and follow-up of glaucoma. *Curr Opinion in Ophthalmol*, 2011; 22: 115-123.
7. Leung CK, et al. Evaluation of retinal nerve fiber layer progression in glaucoma: a comparison between spectral-domain and time-domain optical coherence tomography. *Ophthalmol*, 2011 Aug; 118(8): 1558-62.
8. Leung CK, et al. Retinal nerve fiber layer imaging with spectral-domain optical coherence tomography: patterns of retinal nerve fiber layer progression. *Ophthalmol*, 2012 Sep; 119(9): 1858-66.
9. Leung CK, et al. Retinal nerve fiber layer imaging with spectral-domain optical coherence tomography: a prospective analysis of age-related loss. *Ophthalmol*, 2012 Apr; 119(4): 731-7.

10. Leung CK, et al. Retinal nerve fiber layer imaging with spectral-domain optical coherence tomography: interpreting the RNFL maps in healthy myopic eyes. *Invest Ophthalmol Vis Sci*. 2012 Oct 17; 53(11): 7194-200.
11. Wollstein G, et al. Optical coherence tomography longitudinal evaluation of retinal nerve fiber layer thickness in glaucoma. *Arch Ophthalmol*. 2005 Apr; 123(4): 464-70.
12. Sun KR, et al. Progression detection capability of macular thickness in advanced glaucomatous eyes. *Ophthalmol*. 2012 Feb; 119(2): 308-13.
13. Kim C, Radcliffe N. One year of glaucoma research in review: 2011 to 2012. *Asia-Pacific J of Ophthalmol*, Nov/Dec 2012; 1(6): 364-373.
14. Roh KH, Jeoung JW, Park KH, Yoo BW, Kim DM. Long-term reproducibility of Cirrus HD optical coherence tomography deviation map in clinically stable glaucomatous eyes. *Ophthalmol*, 2013; 120(5): 969-977.
15. Medeiros F et al. Estimating the rate of retinal ganglion cell loss in glaucoma. *Am J Ophthalmol*, 2012; 154: 814-824.
16. Giani A et al. Reproducibility of retinal thickness measurements on normal and pathologic eyes by different optical coherence tomography instruments. *Am J Ophthalmol*, 2010; 150: 815-824.
17. Hirasawa H et al. Peripapillary retinal nerve fiber layer thickness determined by spectral-domain optical coherence tomography in ophthalmologically normal eyes. *Arch Ophthalmol*, 2010; 128(11): 1420-1426.
18. Hong SW, Ahn MD, Kand SH, Im SK. Analysis of peripapillary retinal nerve fiber distribution in normal young adults. *IOVS*, July 2010; 51(7): 3515-3523.
19. Vernon SA, Rotchford AP, et al. Peripapillary retinal nerve fiber layer thickness in highly myopic Caucasians as measured by Stratus optical coherence tomography. *Br J Ophthalmol*, 2008; 92: 1076-1080.
20. Ferreras A, Pablo LE, et al. Mapping standard automated perimetry to the peripapillary retinal nerve fiber layer in glaucoma. *IOVS*, July 2008; 49(7): 3018-3025.
21. <http://www.fda.gov/downloads/MedicalDevices/NewsEvents/WorkshopsConferences/UCM326697.pdf>
22. Savini G, Carbonelli M, Barboni P. Spectral-domain optical coherence tomography for the diagnosis and follow-up of glaucoma. *Curr Opin Ophthalmol*, 2011; 22: 115-123.
23. Mwanza JC, Chang RT, Bunde DL, et al. Reproducibility of peripapillary retinal nerve fiber layer thickness and optic nerve head parameters measured by Cirrus HD-OCT in glaucomatous eyes. *IOVS*, 2010; 51: 5724-5730.
24. Yim SY, Park HL, Park CK. The effects of peripapillary atrophy on the diagnostic ability of Stratus and Cirrus OCT in the analysis of optic nerve head parameters and disc size. *IOVS*, 2012; 53: 4475-4484.
25. Leite ML, Zangwill LM, Weinreb RN, et al. Effect of disease severity on the performance of Cirrus spectral-domain OCT for glaucoma diagnosis. *IOVS*, 2010; 51: 4104-4109.
26. Alasil T, Wang K, Keane P, et al. Analysis of normal retinal nerve fiber layer thickness by age, sex, and race using spectral domain optical coherence tomography. *J Glaucoma*, 2013 Sep; 22(7): 532-41.
27. Lim SC, Singh K, Jampel DH, et al. Optic nerve head and retinal nerve fiber layer analysis; A report by the American Academy of Ophthalmology. *Ophthalmol*, 2007; 114: 1937-1949.
28. Vernon SA, Rotchford AP, Nega A, Ryatt S, Tattersal C. Peripapillary retinal nerve fiber layer thickness in highly myopic Caucasians as measured by Stratus optical coherence tomography. *Br J Ophthalmol*, 2008; 92: 1076-1080.
29. Jeoung JW, Park KH. Comparison of Cirrus OCT and Stratus OCT on the ability to detect localized retinal nerve fiber layer defects in preperimetric glaucoma. *IOVS*, 2010; 51: 938-945.

30. Sommer A, Katz J, Quigley HA. Clinically detectable nerve fiber atrophy precedes the onset of glaucomatous field loss. *Arch Ophthalmol*, 1991; 111: 485-490.
31. Tenkumo K, Hirooka K, Baba T, Nitta E, Sato S, Shiraga F. Evaluation of relationship between retinal nerve fiber layer thickness progression and visual field progression in patients with glaucoma. *Jpn J Ophthalmol*, 2013 Jun 25. [Epub ahead of print]
32. Na JH, Sung KR, Baek S, Lee JY, Kim S. Progression of retinal nerve fiber layer thinning in glaucoma assessed by cirrus optical coherence tomography-guided progression analysis. *Curr Eye Res*, 2013 Mar; 38(3): 386-95.
33. Budenz DL, Anderson DR, Varma R, et al. Determinants of normal retinal nerve fiber layer thickness measured by Stratus OCT. *Ophthalmol*, 2007; 114: 1046-1052.
34. Sung KR, Sun JH, Na JH, Lee JY, Lee Y. Progression detection capability of macular thickness in advanced glaucomatous eyes. *Ophthalmol*, 2012; 119: 308-313.
35. Heijl A, Bengtsson B, Hyman L, Leske MC for the Early Manifest Glaucoma Trial Group. Natural history of open-angle glaucoma. *Ophthalmol*, 2009; 116: 2271-2276.
36. Huang D, et al. for the Advanced Imaging for Glaucoma Study (AIGS) Group. Does optic nerve head size variation affect circumpapillary retinal nerve fiber layer thickness measurement by optical coherence tomography? *IOVS*, 2012; 53: 4990-4997.
37. Jampel HD, Sing K, Lin SC, et al. Assessment of visual function in glaucoma. A report by the American Academy of Ophthalmology. *Ophthalmol*, 2011; 118: 986-1002.
38. Medeiros FA, Lisboa R, Weinreb RN, et al. Retinal ganglion cell count estimates associated with early development of visual field defects in glaucoma. *Ophthalmol*, 2013; 120: 736-744.
39. Grewal DS, Tanna AP. Diagnosis of glaucoma and detection of glaucoma progression using spectral domain optical coherence tomography. *Curr Opin Ophthalmol*, 2013; 24: 150-161.
40. Lisboa R, Leite MT, Zangwill LM, et al. Diagnosing preperimetric glaucoma with spectral domain optical coherence tomography. *Ophthalmol*, 2012; 119 (11): 2261-2269.
41. Mwanza JC, Oakley JD, Budenz DL, Anderson DR for the Cirrus Optical Coherence Tomography Normative Database Study Group. Ability of Cirrus HD-OCT Optic Nerve Head Parameters to Discriminate Normal from Glaucomatous Eyes. *Ophthalmol*, 2011; 118: 241-248.
42. Knight OJ, Chang RT, Feuer WJ, Budenz DL. Comparison of retinal nerve fiber layer measurements using time domain and spectral domain optical coherent tomography. *Ophthalmol*, 2009; 116: 1271-1277.
43. Mwanza JC, Durbin MK, Budenz DL for the Cirrus Optical Coherence Tomography Normative Database Study Group. Interocular symmetry in peripapillary retinal nerve fiber layer thickness measured with the Cirrus HD-OCT in healthy eyes. *Am J Ophthalmol*, 2011; 151: 514-521.
44. Mwanza JC, Durbin MK, Budenz DL, et al. for the Cirrus Optical Coherence Tomography Normative Database Study Group. Macular Ganglion Cell–Inner Plexiform Layer: Automated Detection and Thickness Reproducibility with Spectral Domain–Optical Coherence Tomography in Glaucoma *IOVS*, 2011; 52: 7872-7879.
45. Kotowski J, Wollstein G, Folio L, et al. Clinical use of OCT in assessing glaucoma progression. *Ophthalm Surg Lasers Imaging*, 2011; 42: S6-S14.
46. Na JH, Sung KR, Lee JR, et al. Detection of glaucomatous progression by spectral-domain optical coherence tomography. *Ophthalmol*, 2013: in press.

47. Sung KR, Kim DY, Park SB, Kook MS. Comparison of retinal nerve fiber layer thickness measured by Cirrus HD and Stratus optical coherence tomography. *Ophthalmol*, 2009; 116: 1264-1270.
48. Na JH, Sung KR, Baek S, et al. Detection of glaucomatous progression by assessment of segmented macular thickness data obtained using spectral domain optical coherence tomography. *IOVS*, 2012; 53: 3817-3826.
49. Leung CK, Cheung CL, Weinreb RN, Qui K, et al. Evaluation of retinal nerve fiber layer progression in glaucoma: A study on optical coherence tomography guided progression analysis. *IOVS*, 2010; 51: 217-222.
50. Grewal DS, Sehi M, Paauw J, Greenfield DS for the Advanced Imaging in Glaucoma Study Group (AIGS). Detection of progressive retinal nerve fiber layer thickness loss with optical coherence tomography using 4 criteria for functional progression. *J Glaucoma*, 2012; 21: 214-220.
51. Leung CK, Liu S, Weinreb RN, et al. Evaluation of retinal nerve fiber layer progression in glaucoma; A prospective analysis with neuroretinal rim and visual field progression. *Ophthalmol*, 2011; 118: 1551-1557.
52. Medeiros FA, Lisboa R, Weinreb RN, et al. A combined index of structure and function for staging glaucomatous damage. *Arch Ophthalmol*, 2012; 130(9): 107-1116.
53. Medeiros FA, Zangwill LM, Alencar LM, et al. Detection of glaucoma progression with Stratus OCT retinal nerve fiber layer, optic nerve head, and macular thickness measurements. *IOVS*, 2009; 50: 5741-5748.
54. Leung CK, Cheung CY, Lin D, et al. Longitudinal variability of optic disc and retinal nerve fiber layer measurements. *IOVS*, 2008; 49: 4886-4892.
55. Leung CK, Yu M, Weinreb RN, et al. Retinal nerve fiber layer imaging with spectral-domain optical coherence tomography: interpreting the RNFL maps in healthy myopic eyes. *IOVS*, 2012 Oct 17; 53(11): 7194-7200.
56. Lisboa R, Weinreb RN, Medeiros FA. Combining structure and function to evaluate glaucomatous progression: implications for the design of clinical trials. *Curr Opin in Pharmacology*, 2013; 13: 115-122.
57. Grewal DS, Sehi M, Greenfield DS. Diffuse glaucomatous structural and functional damage in the hemifield without significant pattern loss. *Arch Ophthalmol*, 2009; 127(11): 1442-1448.
58. Polo V, Larrosa JM, Ferreras A, et al. Comparative study of retinal nerve fiber layer thickness in normal eyes, ocular hypertensives, preperimetric glaucoma and glaucomatous subjects. *Annals of Ophthalmol*, 2009; 41(1): 24-30.
59. Leung CK, Lam S, Weinreb RN, et al. Retinal nerve fiber layer imaging with Spectral-domain optical coherence tomography. Analysis of the Retinal Nerve Fiber Layer Map for Glaucoma Detection. *Ophthalmol*, 2010; 117: 1684-1691.
60. Aref AA, Budenz DL. Spectral domain optical coherence tomography for the diagnosis and management of glaucoma. *Ophthalmic Surg Lasers Imaging*, 2010; 41: S15-S27.
61. Knight OJ, Girkin CA, Budenz DL, et al for the Cirrus OCT Normative Database Study Group. Effect of race, age, and axial length on optic nerve head parameters and retinal nerve fiber layer thickness measured by Cirrus HD-OCT. *Arch Ophthalmol*, 2012 Mar; 130(3): 312-8.
62. Wollstein G, Schuman JS, Price LL, et al. Optical Coherence Tomography Longitudinal Evaluation of Retinal Nerve Fiber Layer Thickness in Glaucoma. *Arch Ophthalmol*. 2005; 123: 464-470.
63. Harizman N, Oliveira C, Chiang A, et al. The ISNT rule and differentiation of normal from glaucomatous eyes. *Arch Ophthalmol*, 2006; 124: 1579-1583.

64. Wessel JM, Horn FK, Tornow RP, et al. Longitudinal analysis of progression in glaucoma using spectral-domain optical coherence tomography. *Invest Ophthalmol Vis Sci*, 2013 May 1; 54(5): 3613-20.
65. Shin HY, Park HL, Jung K, et al. Glaucoma diagnostic ability of ganglion cell-inner plexiform layer thickness differs according to the location of visual field loss. *Ophthalmol*, 2013, in press.
66. Wollstein G, Kagemann L, Bilonick R, et al. Retinal nerve fibre layer and visual function loss in glaucoma: the tipping point. *BJO*, 2012; 96: 47-52.
67. Hong JT, Sung KR, Cho JW, et al. Retinal Nerve Fiber Layer Measurement Variability with Spectral Domain Optical Coherence Tomography. *Korean J Ophthalmol*, 2012; 26(1): 32-38.
68. Toscano DA, de Ávila MP, Chalita MRC. Reproducibility of peripapillary retinal nerve fiber layer thickness measurements using Spectral Domain OCT in Brazilian patients. *Arq Bras Oftalmol*, 2012; 75(5): 320-3.
69. Sullivan-Mee M, Ruegg CC, Pensyl D, et al. Diagnostic precision of retinal nerve fiber layer and macular thickness asymmetry parameters for identifying early primary open-angle glaucoma. *Am J Ophthalmol*, 2013; 156: 567-577.
70. Schulze A, Lamparter J, Pfeiffer N, et al. Comparison of laser scanning diagnostic devices for early glaucoma detection. *J Glaucoma*, 2014; in press.
71. Faghihi H, Hajizadeh F, Hashemi H, et al. Agreement of two different spectral domain optical coherence tomography instruments for retinal nerve fiber layer measurements. *J Ophthalmic Vis Res*, 2014; 9(1): 31-37.
72. Asrani S, Essaid L, Alder BD, et al. Artifacts in spectral-domain optical coherence tomography measurements in glaucoma. *JAMA Ophthalmol*, 2014; 132(4): 396-402.
73. Kim YJ, Kang MH, Cho HY, et al. Comparative study of macular ganglion cell complex thickness measured by spectral-domain optical coherence tomography in healthy eyes, eyes with preperimetric glaucoma, and eyes with early glaucoma. *Jpn J Ophthalmol*, 2014; 58: 244-251.
74. Danesh-Meyer HV, Yap J, Frampton C, et al. Differentiation of compressive from glaucomatous optic neuropathy with spectral-domain optical coherence tomography. *Ophthalmol*, 2014; in press.
75. Leung CK. Diagnosing glaucoma progression with optical coherence tomography. *Curr Opin Ophthalmol*, 2014; 25: 104-111.
76. Mwanza JC, Budenz DL, Godfrey DG, et al. Diagnostic performance of optical coherence tomography ganglion cell-inner plexiform layer thickness measurements in early glaucoma. *Ophthalmol*, 2014; 121: 849-854.
77. Pradham ZS, Braganza A, Abraham LK. Does the ISNT rule apply to the retinal nerve fiber layer? *J Glaucoma*, 2014; in press.
78. Iverson Sm, Feuer WJ, Shi W, et al. for the Advanced Imaging in Glaucoma Study Group. Frequency of abnormal retinal nerve fibre layer and ganglion cell layer SDOCT scans in healthy eyes and glaucoma suspects in a prospective longitudinal study. *Br J Ophthalmol*, 2014; 98: 920-925.
79. Shin HY, Park HL, Jung Y, et al. Glaucoma diagnostic accuracy of optical coherence tomography parameters in early glaucoma with different types of optic disc damage. *Ophthalmol*, 2014; in press.
80. Mansouri K, Medeiros FA, Tatham AJ, et al. Evaluation of retinal and choroidal thickness by swept-source optical coherence tomography: repeatability and assessment of artifacts. *Ophthalmol*, 2014; 157: 1022-1032.
81. Field MG, Alasil T, Baniyasi N, et al. Facilitating glaucoma diagnosis with intereye retinal nerve fiber layer asymmetry using spectral-domain optical coherence tomography. *J Glaucoma*, 2014; in press.
82. Spaeth GL, Reddy SC. Imaging of the optic disk in caring for patients with glaucoma: ophthalmoscopy and photography remain the gold standard. *Surv Ophthalmol*, 2014; 59: 454-457.

83. Kotowski J, Wollstein G, Ishikawa H, Schuman JS. Imaging of the optic nerve and retinal nerve fiber layer: an essential part of glaucoma diagnosis and monitoring. *Surv Ophthalmol*, 2014; 59: 458-467.
84. Hwang YH, Jeong YC, Kim HK, Sohn YH. Macular ganglion cell analysis for early detection of glaucoma. *Ophthalmol*, 2014; in press.
85. Sarkar KC, Das P, Pal R, Shaw C. Optical coherence tomographic assessment of retinal nerve fiber layer thickness changes before and after glaucoma filtering surgery. *Oman J Ophthalmol*, 2014; 7: 3-8.
86. Hood DC, Raza AS. On improving the use of OCT imaging for detecting glaucomatous damage. *Br J Ophthalmol*, 2014; 98: ii1-ii9.
87. Lee EJ, Kim T, Kim M, Choi YJ. Peripapillary retinoschisis in glaucomatous eyes. *PLoS ONE* 9(2); e90129.
88. Suh MH, Yoo BW, Kim JY, et al. Quantitative assessment of retinal nerve fiber layer defect depth using spectral-domain optical coherence tomography. *Ophthalmol*, 2014; in press.
89. Na JH, Sung KR, Baek SH, et al. Rates and patterns of macular and circumpapillary retinal nerve fiber layer thinning in preperimetric and perimetric glaucomatous eyes. *J Glaucoma*, 2014; in press.
90. Miki A, Medeiros FA, Weinreb RN, et al. Rates of retinal nerve fiber layer thinning in glaucoma suspect eyes. *Ophthalmol*, 2014; in press.
91. Sigal IA, Wang B, Strouthidis NG, et al. Recent advances in OCT imaging of the lamina cribrosa. *Br J Ophthalmol*, 2014; 98: ii34-ii39.
92. Nadler Z, Wang B, Wollstein G, et al. Repeatability of in vivo 3D lamina cribrosa microarchitecture using adaptive optics spectral domain optical coherence tomography. *Biomed Optics Express*, 2014; 5(4): 1114-1123.
93. Wang B, Nevins J, Nadler Z, et al. Reproducibility of in-vivo OCT measured three-dimensional human lamina cribrosa microarchitecture. *PLoS ONE* 9(4); e95526.
94. Kim GA, Kim JH, Lee JM, Park KS. Reproducibility of peripapillary retinal nerve fiber layer thickness measured by spectral domain optical coherence tomography in pseudophakic eyes. *Korean J Ophthalmol*, 2014; 28(2): 138-149.
95. Rolle T, Dallorto L, Briamonta C, Penna RR. Retinal nerve fibre layer and macular thickness analysis with fourier domain optical coherence tomography in subjects with a positive family history of primary open angle glaucoma. *Br J Ophthalmol*, in press.
96. Kim KE, Kim KH, Jeoung JW, et al. Severity-dependent association between ganglion cell inner plexiform layer thickness and macular mean sensitivity in open-angle glaucoma. *Acta Ophthalmol*, 2014; in press.
97. Yamashita T, Kii Y, Tanaka M, et al. Relationship between supernormal sectors of retinal nerve fibre layer and axial length in normal eyes. *Acta Ophthalmol*, 2014; in press.
98. Smith JP, Woods AD, Bi H, et al. Staging glaucoma using stratus OCT in a U.S. veteran population. *Optom Vis Sci*, 2014; 540-548.
99. Lumbroso B, Rispoli M. Guide to interpreting spectral domain optical coherence tomography. INC, 2009.